

Customer: Emory - HSRB
 Address: 1760 Haygood Drive
 City, State, Zip: Atlanta, GA, 30322
 Ph: 404-852-0765 Contact: Mimi Le

Unit Preventative Maintenance – Centrifuge


Date of Service: 01/30/25 Next Service Due: 01/30/26 Tech: JCD Room # E264
 Customer I.D. IRON MAN SES # 8552 MFR: EPPENDORF
 Model # 5430 R Serial # 5428GH223998
 NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26
 Tachometer Serial # C1958046 Tachometer Calibration Due: 12/17/25

	Pass	Fail	N/A
Unit Level Stable	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Assembly	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Latching Mechanism	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Shocks	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Clean Buckets	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Lubricate Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Rotor Bucket O-Rings	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Electrical Power	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Clean Condenser	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Cooling Capacity	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Brakes	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Checks OK for Use	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

RPMs

Display RPM: 3328 Actual RPM: 3330

Technician's Notes:
LID SHOCK NEEDS REPLACED.



Customer:	Emory - HSRB
Address:	1760 Haygood Drive
City, State, Zip:	Atlanta, GA, 30322
Ph: 404-852-0765	Contact: Mimi Le


Unit Preventative Maintenance – Centrifuge

Date of Service: 01/30/25 Next Service Due: 01/30/26 Tech: JCD Room # E264
 Customer I.D. SPIDER-MAN SES # 8553 MFR: EPPENDORF
 Model # 5702 R Serial # 5703FL416261
 NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26
 Tachometer Serial # C1958046 Tachometer Calibration Due: 12/17/25

	Pass	Fail	N/A
Unit Level Stable	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Assembly	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Latching Mechanism	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Shocks	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Clean Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lubricate Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Rotor Bucket O-Rings	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Electrical Power	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Clean Condenser	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Cooling Capacity	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Brakes	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Checks OK for Use	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

RPMs	
Display RPM: <u>2900</u>	Actual RPM: <u>2876</u>

Technician's Notes:



Customer:	Emory - HSRB
Address:	1760 Haygood Drive
City, State, Zip:	Atlanta, GA, 30322
Ph: 404-852-0765	Contact: Mimi Le

Unit Preventative Maintenance – Centrifuge

Date of Service: 01/30/25 Next Service Due: 01/30/26 Tech: JCD Room # E264

Customer I.D. HULK SES # 8554 MFR: EPPENDORF

Model # 5810 R Serial # 5811GG580544

NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26

Tachometer Serial # C1958046 Tachometer Calibration Due: 12/17/25

	Pass	Fail	N/A
Unit Level Stable	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Assembly	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Latching Mechanism	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Shocks	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Clean Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lubricate Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Rotor Bucket O-Rings	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Electrical Power	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Clean Condenser	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Cooling Capacity	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Brakes	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Checks OK for Use	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

RPMs	
Display RPM: <u>2000</u>	Actual RPM: <u>2003</u>

Technician's Notes:
NEEDS SHOCK REPLACED.



Customer:	Emory - HSRB
Address:	1760 Haygood Drive
City, State, Zip:	Atlanta, GA, 30322
Ph: 404-852-0765	Contact: Mimi Le


Unit Preventative Maintenance – Centrifuge

Date of Service: 01/30/25 Next Service Due: 01/30/26 Tech: JCD Room # E264
 Customer I.D. SCARLET WITCH SES # 9419 MFR: EPPENDORF
 Model # 5702 R Serial # 5703FR016733
 NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26
 Tachometer Serial # C1958046 Tachometer Calibration Due: 12/17/25

	Pass	Fail	N/A
Unit Level Stable	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Assembly	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Latching Mechanism	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Shocks	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Clean Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lubricate Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Rotor Bucket O-Rings	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Electrical Power	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Clean Condenser	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Cooling Capacity	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Brakes	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Checks OK for Use	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

RPMs	
Display RPM: <u>2500</u>	Actual RPM: <u>2520</u>

Technician's Notes:



Customer: CHOA - Center for Advanced Pediatrics
 Address: 2174 N Druid Hills Road NE
 City, State, Zip: Atlanta, GA, 30329
 Ph: 404-852-0765 Contact: Mimi Le

Unit Preventative Maintenance – Centrifuge

Date of Service: 01/30/25 Next Service Due: 01/30/26 Tech: JCD Room # 5126
 Customer I.D. NA SES # 9421 MFR: EPPENDORF
 Model # 5702 R Serial # 5703GG316802
 NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26
 Tachometer Serial # C1958046 Tachometer Calibration Due: 12/17/25

	Pass	Fail	N/A
Unit Level Stable	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Assembly	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Latching Mechanism	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Shocks	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Clean Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lubricate Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Rotor Bucket O-Rings	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Electrical Power	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Clean Condenser	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Cooling Capacity	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Brakes	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Checks OK for Use	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

RPMs

Display RPM: 2900 Actual RPM: 2877

Technician's Notes:



Customer:	Emory - HSRB
Address:	1760 Haygood Drive
City, State, Zip:	Atlanta, GA, 30322
Ph: 404-852-0765	Contact: Mimi Le

Unit Preventative Maintenance – Centrifuge

Date of Service: 01/30/25 Next Service Due: 01/30/26 Tech: JCD Room # E264
 Customer I.D. BLACK WIDOW SES # 10881 MFR: EPENDORF
 Model # 5702 R Serial # 5703FQ216633
 NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26
 Tachometer Serial # C1958046 Tachometer Calibration Due: 12/17/25

	Pass	Fail	N/A
Unit Level Stable	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Assembly	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Latching Mechanism	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lid Shocks	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Clean Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Lubricate Buckets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Rotor Bucket O-Rings	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Electrical Power	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Clean Condenser	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Cooling Capacity	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Brakes	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Checks OK for Use	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

RPMs	
Display RPM: <u>1400</u>	Actual RPM: <u>1397</u>

Technician's Notes:

