

Customer: CHOA - Advanced Pediatrics
Address: 2174 N Druid Hills Road
City, State, Zip: Atlanta, GA, 30329
Ph: 404-852-0765 Contact: Mimi Le
☐ 1st Round ☒ 2nd Round



Unit Preventative Maintenance – Refrigerator or Freezer

Date of Service: 02/27/25 Next Service Due: 08/27/25 Tech: CD Room # 5126
Customer I.D. OKOYE -20C SES # 9424 MFR: VWR
Model # HCUCFS-0420 Serial # VWR-041805280116-1808
NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26

	Pass	Fail	N/A		Pass	Fail	N/A	NEW
Door Seal	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Clear Evaporator	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Gaskets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Drain	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Hinges & Fastening	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Fans	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Handles	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Insulation Lines	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Clean Condenser	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Battery	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Condenser Fan Motor	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	CFM Counter	<u>NA</u>			

Temperature

Setpoint: NA Display: NA NIST: -20.3C Chart: NA Adjust: NA

Compressor

Voltage: 115 Amperage: 0.5 Model: B43C5B

☐ Inverter Normal ☐ Inverter Error (See Notes) ☒ Inverter N/A

1 Unit Total Amperage NA High Alarm Set NA Battery Charge Circuit
0.9 Amps Rated NA Low Alarm Set GOOD Wiring Connections
115 Volts Rated

Technician's Notes:

UNIT NEEDS TO BE DEFROSTED - ALMOST THE ENTIRE CHAMBER IS ICED OVER. HAD TO REMOVE ICE FROM THE UNIT TO BE ABLE TO CLOSE THE DOOR. UNABLE TO SEE THE SETTING ON THE DIAL CONTROL (LAST ROUND IT WAS SET TO MAX COLD)

Customer: CHOA - Advanced Pediatrics
 Address: 2174 N Druid Hills Road
 City, State, Zip: Atlanta, GA, 30329
 Ph: 404-852-0765 Contact: Mimi Le
☐ 1st Round ☒ 2nd Round



Unit Preventative Maintenance – Refrigerator or Freezer

Date of Service: 02/27/25 Next Service Due: 08/27/25 Tech: CD Room # 5126
 Customer I.D. SHURI 4C SES # 9425 MFR: VWR
 Model # HCUCFS-0404 Serial # VWR-051808050362-1809
 NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26

	Pass	Fail	N/A		Pass	Fail	N/A	NEW
Door Seal	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Clear Evaporator	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Gaskets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Drain	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Hinges & Fastening	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Fans	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Handles	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Insulation Lines	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Clean Condenser	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Battery	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Condenser Fan Motor	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	CFM Counter	<u>NA</u>			

Temperature

Setpoint: DIAL 2 Display: NA NIST: 3.1C Chart: NA Adjust: NA

Compressor

Voltage: 114 Amperage: 0.5 Model: EYA40U6L

☐ Inverter Normal ☐ Inverter Error (See Notes) ☒ Inverter N/A

1 Unit Total Amperage NA High Alarm Set NA Battery Charge Circuit
0.8 Amps Rated NA Low Alarm Set GOOD Wiring Connections
115 Volts Rated

Technician's Notes:

UNIT HAS A DIAL CONTROL

Customer: Chidren's Healthcare of Atlanta - CAP
 Address: 2174 N Druid Hills Road
 City, State, Zip: Atlanta, GA, 30329
 Ph: 404-852-0765 Contact: Mimi Le
☒ 1st Round ☐ 2nd Round



Unit Preventative Maintenance – Refrigerator or Freezer

Date of Service: 09/04/25 Next Service Due: 03/04/26 Tech: CD Room # 5126
 Customer I.D. OKOYE SES # 9424 MFR: VWR
 Model # HCUCFS-0420 Serial # VWR-041805280116-1808
 NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26

	Pass	Fail	N/A		Pass	Fail	N/A	NEW
Door Seal	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Clear Evaporator	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Gaskets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Drain	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Hinges & Fastening	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Fans	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Handles	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Insulation Lines	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Clean Condenser	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Battery	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Condenser Fan Motor	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	CFM Counter	<u>NA</u>			

Temperature

Setpoint: DIAL 7 Display: NA NIST: -26.4C Chart: NA Adjust: NA

Compressor

Voltage: 115 Amperage: 0.5 Model: B43C5B

☐ Inverter Normal ☐ Inverter Error (See Notes) ☒ Inverter N/A

1 Unit Total Amperage NA High Alarm Set NA Battery Charge Circuit
0.9 Amps Rated NA Low Alarm Set GOOD Wiring Connections
115 Volts Rated

Technician's Notes:

UNIT IS SET TO MAX COLD POSITON ON THE DIAL

Customer: Chidren's Healthcare of Atlanta - CAP
 Address: 2174 N Druid Hills Road
 City, State, Zip: Atlanta, GA, 30329
 Ph: 404-852-0765 Contact: Mimi Le
☒ 1st Round ☐ 2nd Round



Unit Preventative Maintenance – Refrigerator or Freezer

Date of Service: 09/04/25 Next Service Due: 03/04/26 Tech: CD Room # 5126
 Customer I.D. SHURI SES # 9425 MFR: VWR
 Model # HCUCFS-0404 Serial # VWR-051808050362-1809
 NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26

	Pass	Fail	N/A		Pass	Fail	N/A	NEW
Door Seal	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Clear Evaporator	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Gaskets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Drain	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Hinges & Fastening	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Fans	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Handles	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Insulation Lines	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Clean Condenser	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Battery	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Condenser Fan Motor	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	CFM Counter	<u>NA</u>			

Temperature

Setpoint: DIAL 2 Display: NA NIST: 4.1C Chart: NA Adjust: NA

Compressor

Voltage: 115 Amperage: 0.5 Model: EYA40U6L

☐ Inverter Normal ☐ Inverter Error (See Notes) ☒ Inverter N/A

1 Unit Total Amperage NA High Alarm Set NA Battery Charge Circuit
0.8 Amps Rated NA Low Alarm Set GOOD Wiring Connections
115 Volts Rated

Technician's Notes:
