



Customer: Emory - HSRB  
Address: 1760 Haygood Dr  
City, State, Zip: Atlanta GA 30332  
Ph: 404-852-0765 Contact: Mimi Le

## Unit Preventative Maintenance – Centrifuge

Date of Service: 01/12/26 Next Service Due: 01/12/27 Tech: AB Room # E264

Customer I.D. Iron Man SES # 8552 MFR: Eppendorf

Model # 5430R Serial # 5428GH223998

NIST Meter Serial # NA NIST Meter Calibration Due:

Tachometer Serial # 1859755 Tachometer Calibration Due: 03/31/26

|                      | Pass                                | Fail                                | N/A                                 |
|----------------------|-------------------------------------|-------------------------------------|-------------------------------------|
| Unit Level Stable    | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| Lid Assembly         | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| Latching Mechanism   | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| Lid Shocks           | <input type="checkbox"/>            | <input checked="" type="checkbox"/> | <input type="checkbox"/>            |
| Clean Buckets        | <input type="checkbox"/>            | <input type="checkbox"/>            | <input checked="" type="checkbox"/> |
| Lubricate Buckets    | <input type="checkbox"/>            | <input type="checkbox"/>            | <input checked="" type="checkbox"/> |
| Rotor Bucket O-Rings | <input type="checkbox"/>            | <input type="checkbox"/>            | <input checked="" type="checkbox"/> |
| Electrical Power     | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| Clean Condenser      | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| Cooling Capacity     | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| Brakes               | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| Checks OK for Use    | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |

### RPMs

Setpoint RPM: 3692 Display RPM: 3692 Actual RPM: 3694

Technician's Notes:

Unit needs a new lid shock.



Customer: Emory - HSRB  
 Address: 1760 Haygood Dr  
 City, State, Zip: Atlanta GA 30332  
 Ph: 404-852-0765 Contact: Mimi Le

## Unit Preventative Maintenance – Centrifuge

Date of Service: 01/12/26 Next Service Due: 01/12/27 Tech: AB Room # E264

Customer I.D. Spider Man SES # 8553 MFR: Eppendorf

Model # 5702R Serial # 5703FL416261

NIST Meter Serial # NA NIST Meter Calibration Due:

Tachometer Serial # 1859755 Tachometer Calibration Due: 03/31/26

|                      | Pass                                | Fail                     | N/A                                 |
|----------------------|-------------------------------------|--------------------------|-------------------------------------|
| Unit Level Stable    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lid Assembly         | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Latching Mechanism   | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lid Shocks           | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Clean Buckets        | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lubricate Buckets    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Rotor Bucket O-Rings | <input type="checkbox"/>            | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Electrical Power     | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Clean Condenser      | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Cooling Capacity     | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Brakes               | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Checks OK for Use    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |

### RPMs

Setpoint RPM: 3600 Display RPM: 3.6 Actual RPM: 3598

Technician's Notes:

---



---



---



Customer: Emory - HSRB  
Address: 1760 Haygood Dr  
City, State, Zip: Atlanta GA 30332  
Ph: 404-852-0765 Contact: Mimi Le

## Unit Preventative Maintenance – Centrifuge

Date of Service: 01/12/26 Next Service Due: 01/12/27 Tech: AB Room # E264

Customer I.D. The Hulk SES # 8554 MFR: Eppendorf

Model # 5810R Serial # 5811GG580544

NIST Meter Serial # NA NIST Meter Calibration Due:

Tachometer Serial # 1859755 Tachometer Calibration Due: 03/31/26

|                      | Pass                                | Fail                     | N/A                                 |
|----------------------|-------------------------------------|--------------------------|-------------------------------------|
| Unit Level Stable    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lid Assembly         | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Latching Mechanism   | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lid Shocks           | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Clean Buckets        | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lubricate Buckets    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Rotor Bucket O-Rings | <input type="checkbox"/>            | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Electrical Power     | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Clean Condenser      | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Cooling Capacity     | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Brakes               | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Checks OK for Use    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |

### RPMs

Setpoint RPM: 2000 Display RPM: 2000 Actual RPM: 2000

Technician's Notes:

---

---

---



Customer: Emory - HSRB  
 Address: 1760 Haygood Dr  
 City, State, Zip: Atlanta GA 30332  
 Ph: 404-852-0765 Contact: Mimi Le

## Unit Preventative Maintenance – Centrifuge

Date of Service: 01/12/26 Next Service Due: 01/12/27 Tech: AB Room # E264

Customer I.D. Scarlet Witch SES # 9419 MFR: Eppendorf

Model # 5702R Serial # 5703FR016733

NIST Meter Serial # NA NIST Meter Calibration Due:

Tachometer Serial # 1859755 Tachometer Calibration Due: 03/31/26

|                      | Pass                                | Fail                     | N/A                                 |
|----------------------|-------------------------------------|--------------------------|-------------------------------------|
| Unit Level Stable    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lid Assembly         | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Latching Mechanism   | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lid Shocks           | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Clean Buckets        | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lubricate Buckets    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Rotor Bucket O-Rings | <input type="checkbox"/>            | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Electrical Power     | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Clean Condenser      | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Cooling Capacity     | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Brakes               | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Checks OK for Use    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |

### RPMs

Setpoint RPM: 2500 Display RPM: 2.5 Actual RPM: 2514

Technician's Notes:


---



---




---



Date of Service: 01/12/26 Next Service Due: 01/12/27 Tech: AB Room # 5126  
Customer I.D. NA SES # 9421 MFR: Eppendorf  
Model # 5720R Serial # 5703GG316802  
NIST Meter Serial # NA NIST Meter Calibration Due: \_\_\_\_\_  
Tachometer Serial # 1859755 Tachometer Calibration Due: 03/31/26

Setpoint RPM: 3100      Display RPM: 3.1      Actual RPM: 3096

● 102 White Park Drive - Building 500 ● Dallas, GA 30132 ● (770) 426-6678 ● [www.southeastscientific.com](http://www.southeastscientific.com) ●



Customer: Emory - HSRB  
Address: 1760 Haygood Dr  
City, State, Zip: Atlanta GA 30332  
Ph: 404-852-0765 Contact: Mimi Le

## Unit Preventative Maintenance – Centrifuge

Date of Service: 01/12/26 Next Service Due: 01/12/27 Tech: AB Room # E264

Customer I.D. Black Widow SES # 10881 MFR: Eppendorf

Model # 5702R Serial # 5703FQ216633

NIST Meter Serial # NA NIST Meter Calibration Due:

Tachometer Serial # 1859755 Tachometer Calibration Due: 03/31/26

|                      | Pass                                | Fail                     | N/A                                 |
|----------------------|-------------------------------------|--------------------------|-------------------------------------|
| Unit Level Stable    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lid Assembly         | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Latching Mechanism   | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lid Shocks           | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Clean Buckets        | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Lubricate Buckets    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Rotor Bucket O-Rings | <input type="checkbox"/>            | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Electrical Power     | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Clean Condenser      | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Cooling Capacity     | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Brakes               | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |
| Checks OK for Use    | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/>            |

### RPMs

Setpoint RPM: 1400 Display RPM: 1.4 Actual RPM: 1391

Technician's Notes:

---

---

---