

Customer: CHOA - Center for Advanced Pediatrics
 Address: 2174 N Druid Hills Road NE
 City, State, Zip: Atlanta, GA, 30329
 Ph: 404-852-0765 Contact: Mimi Le
 1st Round 2nd Round



Unit Preventative Maintenance – Refrigerator or Freezer

Date of Service: 03/10/26 Next Service Due: 09/10/26 Tech: CD Room # 5126
 Customer I.D. OKOYE SES # 9424 MFR: VWR
 Model # HCUCFS-0420 Serial # VWR-041805280116-1808
 NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26

	Pass	Fail	N/A		Pass	Fail	N/A	NEW
Door Seal	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Clear Evaporator	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Gaskets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Drain	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Hinges & Fastening	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Fans	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Handles	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Insulation Lines	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Clean Condenser	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Battery	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Condenser Fan Motor	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	CFM Counter	<u>NA</u>			

Temperature

Setpoint: DIAL 7 Display: NA NIST: -18.2C Chart: NA Adjust: NA

Compressor

Voltage: 116 Amperage: 0.5 Model: B43C5B

Inverter Normal Inverter Error (See Notes) Inverter N/A

1 Unit Total Amperage NA High Alarm Set NA Battery Charge Circuit
0.9 Amps Rated NA Low Alarm Set GOOD Wiring Connections
115 Volts Rated

Technician's Notes:

Customer: CHOA - Center for Advanced Pediatrics
 Address: 2174 N Druid Hills Road NE
 City, State, Zip: Atlanta, GA, 30329
 Ph: 404-852-0765 Contact: Mimi Le
 1st Round 2nd Round



Unit Preventative Maintenance – Refrigerator or Freezer

Date of Service: 03/10/26 Next Service Due: 09/10/26 Tech: CD Room # 5126
 Customer I.D. SHURI SES # 9425 MFR: VWR
 Model # HCUCFS-0404 Serial # VWR-051808050362-1809
 NIST Meter Serial # T-281182 NIST Meter Calibration Due: 06/26/26

	Pass	Fail	N/A		Pass	Fail	N/A	NEW
Door Seal	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Clear Evaporator	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Gaskets	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Drain	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Hinges & Fastening	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Evaporator Fans	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Handles	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Insulation Lines	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Clean Condenser	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Battery	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Condenser Fan Motor	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	CFM Counter	<u>NA</u>			

Temperature

Setpoint: DIAL 2 Display: NA NIST: 3.8C Chart: NA Adjust: NA

Compressor

Voltage: 116 Amperage: 0.5 Model: EYA40U6L

Inverter Normal Inverter Error (See Notes) Inverter N/A

1 Unit Total Amperage NA High Alarm Set NA Battery Charge Circuit
0.8 Amps Rated NA Low Alarm Set GOOD Wiring Connections
115 Volts Rated

Technician's Notes:

