



Accutech Calibrations
50 Church St
Rockmart, GA 30153
info@accutechcalibrations.com
770-715-2876

INVOICE/WORKSHEET

Building: HSRB - Emory	Invoice #: 010526HT
Address: Haygood	PO/Acct. #:
Client Name: CTDC	Phone #: 404-727-2343
Service Date: January 5, 2026	Page #: 1 of 1
Lab/Room #: E263	Email: uyen.quynh.nguyen.le@emory.edu

Calibration Record

		Sample Draws					Final Draws			
PIPETTE NUMBER	VOLUME (μL)	LOW	MID	HIGH	PASS/FAIL	SERVICE REQUIRED	LOW	MID	HIGH	COST
1	RU12899	10	Out of Tolerance			Fail	Calibrate	1.00	5.00	10.01
2	OU17272	2	Out of Tolerance			Fail	Calibrate	0.50	1.00	2.00
3	RU10797	10	Out of Tolerance			Fail	Calibrate	1.00	5.01	10.00
4	SU11623	20	2.00	10.01	20.04	Pass	OK			
5	SU11612	20	Out of Tolerance			Fail	Calibrate	2.10	10.00	20.01
6	OU14246	100	10.10	49.80	100.32	Pass	OK			
7	RU12909	100	10.00	49.85	99.96	Pass	OK			
8	SU20765	100	10.00	49.54	100.07	Pass	OK			
9	PU22457	100	10.04	49.77	100.03	Pass	OK			
10	RU12914	100	Out of Tolerance			Fail	Calibrate	10.01	49.83	99.68
11	QU37783	200	20.00	99.23	199.57	Pass	OK			
12	OU17222	200	Below Standards			Fail	Calibrate	20.43	99.87	200.06
13	RU04927	300	Below Standards			Fail	Calibrate	29.96	149.88	301.02
14	SU11679	300	Out of Tolerance			Fail	Calibrate	29.45	149.57	300.39
15	RU05037	300	29.54	148.93	301.80	Pass	OK			
16	SU11674	300	Out of Tolerance			Fail	Calibrate	29.73	150.68	301.55
17	OU17090	12-CH 300	Below Standards			Fail	Calibrate	30.21	149.89	300.78
18	SU11507	1000	Out of Tolerance			Fail	Calibrate	198.87	500.65	1002.39
19	PU22492	1000	Out of Tolerance			Fail	Calibrate	199.82	149.97	1003.65
20	SU04580	1000	199.66	498.87	1004.42	Pass	OK			
21	RU12929	1000	201.25	501.44	1003.96	Pass	OK			
22	SU11505	1000	Out of Tolerance			Fail	Calibrate	201.33	500.57	1002.21
23	QU37735	1000	200.38	498.99	1000.71	Pass	OK			
24	RU12924	1000	Out of Tolerance			Fail	Calibrate	201.95	500.37	1002.59
25										
26										
27										
28										
29										
30										
31										
32										
33										
34										
35										
36										
37										
38										
39										
40										
41										
42										
43										
44										
45										
46										
47										
48										
49										
50										

Page Total:

Parts Replaced:	Each
	Each
	Each
	Each

TOTAL AMOUNT DUE*:

TECHNICIAN: Heather Thompson

CUSTOMER SIGNATURE: _____

Balance: 11909571

Nist Wt.: 2069

Water C°: 20

Z-Factor: 1.0018

Calibrated: 14 of 24

NEXT CALIBRATION DUE DATE: July 2026