



Accutech Calibrations  
 1219 Hardin Bridge Road  
 Euharlee, GA 30145  
[info@accutechcalibrations.com](mailto:info@accutechcalibrations.com)  
 770-715-2876

**INVOICE/WORKSHEET**

Building: HSRB - Emory	Invoice #: 010725ZC-2
Address: Haygood	PO/Acct. #:
Client Name: CTDC	Phone #: 727-2343
Service Date: January 7, 2025	Page #: 1 of 1
Lab/Room #: E264	Email: <a href="mailto:uyen.quynh.nguyen.le@emory.edu">uyen.quynh.nguyen.le@emory.edu</a>

**Calibration Record**

LINE	PIPETTE NUMBER	VOLUME	Sample Draws				PASS/FAIL	SERVICE REQUIRED	Final Draws			COST
			LOW	MID	HIGH	LOW			MID	HIGH		
1	RU12924	1000	197.98	498.85	1002.75	Pass	Ok				\$0.00	
2	SU11507	1000	Below Standards			Fail	Calibrate	200.50	500.81	1002.49	\$25.00	
3	QU37735	1000	198.65	499.81	1003.37	Pass	Ok				\$0.00	
4	SU11679	300	49.21	149.85	300.51	Pass	Ok				\$0.00	
5	OU17222	200	Out of Tolerance			Fail	Calibrate	49.88	99.67	200.31	\$25.00	
6	QU37783	200	49.99	100.01	200.21	Pass	Ok				\$0.00	
7	RU12909	100	20.01	50.03	100.06	Pass	Ok				\$0.00	
8	SU20765	100	20.01	50.03	100.06	Pass	Ok				\$0.00	
9	RU12914	100	Below Standards			Fail	Calibrate	20.00	50.06	100.11	\$25.00	
10	SU11612	20	5.00	9.99	19.97	Pass	Ok				\$0.00	
11	RU10797	10	Out of Tolerance			Fail	Calibrate	1.98	4.99	10.01	\$25.00	
12	RU12899	10	Out of Tolerance			Fail	Calibrate	2.00	5.00	9.99	\$25.00	
13	OU17090	12CH 300	Below Standards			Fail	Calibrate	50.31	150.11	300.03	\$35.00	
14	SU11508	1000	199.21	500.11	1003.27	Pass	Ok				\$0.00	
15	SU04580	1000	197.69	498.93	1000.21	Pass	Ok				\$0.00	
16	RU04927	300	Below Standards			Fail	Calibrate	50.01	150.16	301.03	\$25.00	
17	OU17272	2	0.50	1.00	2.00	Pass	Ok				\$0.00	
18	SU11674	300	Out of Tolerance			Fail	Calibrate	49.35	149.69	300.51	\$25.00	
19	OU14246	100	20.01	49.99	100.06	Pass	Ok				\$0.00	
20	SU19322	100	19.98	49.89	100.03	Pass	Ok				\$0.00	
21	SU11623	20	5.00	10.01	19.99	Pass	Ok				\$0.00	
22	L1436307T	1000	Below Standards			Fail	Calibrate	200.01	499.99	998.79	\$25.00	
23	K1429559T	200	49.87	99.95	200.53	Pass	Ok				\$0.00	
24	L1435505T	100	19.90	49.96	100.21	Pass	Ok				\$0.00	
25	K1429343T	100	20.10	50.06	99.98	Pass	Ok				\$0.00	
26	L1434323T	20	Below Standards			Fail	Calibrate	5.00	10.01	20.03	\$25.00	
27	L1436089T	2	0.50	1.00	2.00	Pass	Ok				\$0.00	
28	SU11505	1000	Out of Tolerance			Fail	Calibrate	197.75	496.01	1006.93	\$25.00	
29	RU12929	1000	Out of Tolerance			Fail	Calibrate	198.11	495.59	1005.70	\$25.00	
30	RU05037	300	49.56	150.01	300.53	Pass	Ok				\$0.00	
31	SU11611	20	4.91	9.96	20.07	Pass	Ok				\$0.00	
32												
33												
34												
35												
36												
37												
38												
39												
40												
41												
42												
43												
44												
45												
46												
47												
48												
49												
50												

Page Total: \$310.00

Parts Replaced:	Each
	Each
	Each
	Each

**TOTAL AMOUNT DUE\*: \$310.00**

TECHNICIAN: Zachary Cleaver

CUSTOMER SIGNATURE: \_\_\_\_\_

Balance: 11909571

Nist Wt.: 2069

Water C°: 22

Z-Factor: 1.0022

# Calibrated: **13 of 31**

NEXT CALIBRATION DUE DATE: JULY 2025